



CERTIFICATE



This is to certify that

The laboratory SANA Srl accredited by ACCREDIA with N° 0747 according to the UNI EN ISO 17025:2018 has performed

*a test (AFNOR - NF T72-110)
using bactericidal activity, including Staphylococcus aureus ATCC 6538
from 5th to 12th October 2023*

*with STEAM GENERATOR "EV" of the company EV INTERNATIONAL Ltd.
in accordance with AFNOR - NF T72-110 STANDARD for Steam Disinfection Device (SDD)*

*Quantitative surface test with mechanical action for the evaluation
of bactericidal activity of dry steam used in the medical area.*

*This certificate attests to the efficacy of the Steam Disinfection Device under specified procedure.
It confirms the efficacy result measured > 99,99% in clean conditions of the viability of the counts
under examination for bactericidal activity, endorsing its suitability for disinfection purposes.*



*Dr. Marta Colatosti
Laboratory director*



signature and stamp